



CERTIFICATE OF ANALYSIS

STP Prime CCS Comp Cal Cartridge
Lot Number 17006023

Ref Number 53105
Exp Date 2017-12-15
Date of manufacture 2016-12-15

Cal A	L/N	16354004		Cal B	L/N	16354005		Flush	L/N	16350017	
		Spec	Pass/Fail			Spec	Pass/Fail			Spec	Pass/Fail
	pH	7.350	N/A		pH	6.837-6.843	Pass		pCO2 mmHg	51.6	N/A
	pCO2 mmHg	24.6	N/A		Na mmol/L	68.0-69.0	Pass		pO2 mmHg	101.9	N/A
	Na mmol/L	132.1-133.9	Pass		K mmol/L	9.92-10.08	Pass				
	K mmol/L	3.97-4.03	Pass		Cl mmol/L	45.6-46.4	Pass				
	Cl mmol/L	105.2-106.8	Pass		iCa mmol/L	2.17-2.23	Pass				
	iCa mmol/L	1.08-1.12	Pass		Glu mg/dL	197-203	Pass				
	Glu mg/dL	79-81	Pass		Lac mmol/L	9.5-10.5	Pass				
	Lac mmol/L	1.8-2.2	Pass								

Reference L/N 16350004

	Spec	Pass/Fail
KCl	1.8-2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.


Approval

QC Supervisor
Title

1/30/2017
Date