



CERTIFICATE OF ANALYSIS

STP Prime CCS Comp Cal Cartridge
Lot Number 16222061

Ref Number 52861
Exp Date 2017-07-19
Date of Manufacture 2016-07-19

Cal A L/N	16201032		Cal B L/N	16201033		Flush L/N	16204002	
	Spec	Pass/Fail		Spec	Pass/Fail		Spec	Pass/Fail
pH	7.353	N/A	pH	6.837-6.843	Pass	pCO2 mmHg	49.9	N/A
pCO2 mmHg	24.2	N/A	Na mmol/L	68.0-69.0	Pass	pO2 mmHg	102.0	N/A
Na mmol/L	132.1-133.9	Pass	K mmol/L	9.92-10.08	Pass			
K mmol/L	3.97-4.03	Pass	Cl mmol/L	45.6-46.4	Pass			
Cl mmol/L	105.2-106.8	Pass	iCa mmol/L	2.17-2.23	Pass			
iCa mmol/L	1.08-1.12	Pass	Glu mg/dL	197-203	Pass			
Glu mg/dL	79-81	Pass	Lac mmol/L	9.5-10.5	Pass			
Lac mmol/L	1.8-2.2	Pass						

Reference L/N 16201031

	Spec	Pass/Fail
KCl	1.8-2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Approval

QC Supervisor

Title

8/29/2016

Date