



CERTIFICATE OF ANALYSIS

STP Prime CCS Comp Cal Cartridge
Lot Number 16323025

Ref Number 52861
Exp Date 2017-09-28
Date of Manufacture 2016-09-28

Cal A	L/N	16272031		Cal B	L/N	16272033		Flush	L/N	16301037	
		Spec	Pass/Fail			Spec	Pass/Fail			Spec	Pass/Fail
	pH	7.339	N/A		pH	6.837-6.843	Pass		pCO2 mmHg	51.2	N/A
pCO2	mmHg	25.6	N/A	Na	mmol/L	68.0-69.0	Pass		pO2 mmHg	102.0	N/A
Na	mmol/L	132.1-133.9	Pass	K	mmol/L	9.92-10.08	Pass				
K	mmol/L	3.97-4.03	Pass	Cl	mmol/L	45.6-46.4	Pass				
Cl	mmol/L	105.2-106.8	Pass	iCa	mmol/L	2.17-2.23	Pass				
iCa	mmol/L	1.08-1.12	Pass	Glu	mg/dL	197-203	Pass				
Glu	mg/dL	79-81	Pass	Lac	mmol/L	9.5-10.5	Pass				
Lac	mmol/L	1.8-2.2	Pass								

Reference L/N 16272034

	Spec	Pass/Fail
KCl	1.8-2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.


Approval

QC Supervisor
Title

1/13/2017
Date