



CERTIFICATE OF ANALYSIS

STP Prime CCS Comp Cal Cartridge
Lot Number 16323025

Ref Number 53469
Exp Date 2017-09-28
Date of Manufacture 2016-09-28

Cal A L/N	16272031		Cal B L/N	16272033		Flush L/N	16301037	
	Spec	Pass/Fail		Spec	Pass/Fail		Spec	Pass/Fail
pH	7.339	N/A	pH	6.837-6.843	Pass	pCO2 mmHg	51.2	N/A
pCO2 mmHg	25.6	N/A	Na mmol/L	68.0-69.0	Pass	pO2 mmHg	102.0	N/A
Na mmol/L	132.1-133.9	Pass	K mmol/L	9.92-10.08	Pass			
K mmol/L	3.97-4.03	Pass	Cl mmol/L	45.6-46.4	Pass			
Cl mmol/L	105.2-106.8	Pass	iCa mmol/L	2.17-2.23	Pass			
iCa mmol/L	1.08-1.12	Pass	Glu mg/dL	197-203	Pass			
Glu mg/dL	79-81	Pass	Lac mmol/L	9.5-10.5	Pass			
Lac mmol/L	1.8-2.2	Pass						

Reference L/N 16272034

	Spec	Pass/Fail
KCl	1.8-2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Approval

QC Supervisor

Title

12/28/2016

Date