



CERTIFICATE OF ANALYSIS

STP Prime Cal Cartridge ABG
Lot Number 16222061

Ref Number 53104
Exp Date 2017-07-19
Date of Manufacture 2016-07-19

Cal A	L/N	16201032		Cal B	L/N	16201033		Flush	L/N	16204002	
		Spec	Pass/Fail			Spec	Pass/Fail			Spec	Pass/Fail
	pH	7.353	N/A		pH	6.837-6.843	Pass		pCO2 mmHg	49.9	N/A
	pCO2 mmHg	24.2	N/A						pO2 mmHg	102.0	N/A
				Reference	L/N	16201031					
						Spec	Pass/Fail				
				KCl		1.8-2.2	Pass				

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Approval

QC Supervisor
Title

8/22/2016
Date