



CERTIFICATE OF ANALYSIS

STP Prime Cal Cartridge ABG
Lot Number 17006023

Ref Number 53104
Exp Date 2017-12-15
Date of Manufacture 2016-12-15

Cal A	L/N	16354004		Cal B	L/N	16354005		Flush	L/N	16350017	
		Spec	Pass/Fail			Spec	Pass/Fail			Spec	Pass/Fail
	pH	7.350	N/A		pH	6.837-6.843	Pass		pCO2 mmHg	51.6	N/A
	pCO2 mmHg	24.6	N/A						pO2 mmHg	101.9	N/A
				Reference	L/N	16350004					
						Spec	Pass/Fail				
				KCI		1.8-2.2	Pass				

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Approval

QC Supervisor

Title

1/30/2017

Date