



CERTIFICATE OF ANALYSIS

STP Prime Cal Cartridge ABG
Lot Number 16301045

Ref Number 53359
Exp Date 2017-09-23
Date of Manufacture 2016-09-23

Cal A	L/N	16272031		Cal B	L/N	16272033		Flush	L/N	16267082	
		Spec	Pass/Fail			Spec	Pass/Fail			Spec	Pass/Fail
	pH	7.339	N/A		pH	6.837-6.843	Pass		pCO2 mmHg	50.9	N/A
	pCO2 mmHg	25.6	N/A						pO2 mmHg	103.9	N/A
				Reference	L/N	16272034					
						Spec	Pass/Fail				
				KCI		1.8-2.2	Pass				

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Approval

QC Supervisor
Title

11/15/2016
Date