



CERTIFICATE OF ANALYSIS

STP Prime Cal Cartridge ABG
Lot Number 16323025

Ref Number 53360
Exp Date 2017-09-28
Date of Manufacture 2016-09-28

Cal A	L/N	16272031		Cal B	L/N	16272033		Flush	L/N	16301037	
		Spec	Pass/Fail			Spec	Pass/Fail			Spec	Pass/Fail
	pH	7.339	N/A		pH	6.837-6.843	Pass		pCO2 mmHg	51.2	N/A
pCO2 mmHg		25.6	N/A						pO2 mmHg	102.0	N/A
				Reference	L/N	16272034					
						Spec	Pass/Fail				
				KCI		1.8-2.2	Pass				

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Approval

QC Supervisor
Title

1/13/2017
Date