



CERTIFICATE OF ANALYSIS

STP Prime Cal Cartridge ABG
Lot Number 16354046

Ref Number 53360
Exp Date 2017-10-26
Date of Manufacture 2016-10-26

Cal A	L/N	16300042		Cal B	L/N	16300034		Flush	L/N	16316042	
		Spec	Pass/Fail			Spec	Pass/Fail			Spec	Pass/Fail
	pH	7.353	N/A		pH	6.837-6.843	Pass		pCO2 mmHg	51.6	N/A
	pCO2 mmHg	24.6	N/A						pO2 mmHg	101.8	N/A
				Reference	L/N	16300021					
						Spec	Pass/Fail				
				KCI		1.8-2.2	Pass				

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Approval

QC Supervisor
Title

1/10/2017
Date