



CERTIFICATE OF ANALYSIS

Nipro Glucose and β -Ketone Control Solution Level 3
Lot Number 0422105303

Ref Number 11-117
Exp Date 2024.10.31

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	275 - 325	312
Ketone mmol/L	2.5 - 3.5	3.4

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in black ink, appearing to read "Anishi M".

Approval

QC Inspector
Title

5/17/2022
Date