



CERTIFICATE OF ANALYSIS

STP Prime Cal Cartridge ABG
Lot Number 16323025

Ref Number 53463
Exp Date 2017-09-28
Date of Manufacture 2016-09-28

Cal A	L/N	16272031	Cal B	L/N	16272033	Flush	L/N	16301037		
		Spec	Pass/Fail		Spec	Pass/Fail		Spec	Pass/Fail	
	pH	7.339	N/A		pH	6.837-6.843	Pass	pCO2 mmHg	51.2	N/A
	pCO2 mmHg	25.6	N/A					pO2 mmHg	102.0	N/A
Reference L/N 16272034										
					Spec	Pass/Fail				
	KCl				1.8-2.2	Pass				

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Approval

QC Supervisor
Title

12/28/2016
Date