

CERTIFICATE OF ANALYSIS

STP Prime Cal Cartridge ABG Lot Number 17006023 Ref Number 53463

Pass/Fail

Exp Date 2017-12-15

Date of Manufacture 2016-12-15

 Cal A
 L/N
 16354004
 Cal B
 L/N
 16354005
 Flush
 L/N
 16350017

 Spec
 Pass/Fail
 Spec
 Pass/Fail
 Spec
 Spec

pH 7.350 N/A pH 6.837-6.843 Pass pCO2 mmHg 51.6 N/A pCO2 mmHg 24.6 N/A PO2 mmHg 101.9 N/A

Reference L/N 16350004

Spec Pass/Fail KCl 1.8-2.2 Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CRF-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Joyce anasa Approval

QC Supervisor

Title

1/30/2017 Date