



CERTIFICATE OF ANALYSIS

EGB Stat Calibrator Cal Cartridge, 820/1020 Menu
Lot Number 16254001

Ref Number REG00022
Exp Date 2017-08

Cal A	L/N	16230048	Cal B	L/N	16231091	Flush	L/N	16231092
	Spec	Pass/Fail		Spec	Pass/Fail		Spec	Pass/Fail
pH	7.340	N/A	pH	6.837-6.843	Pass	pCO2 mmHg	49.8	N/A
pCO2 mmHg	25.1	N/A	Na mmol/L	68.0-69.0	Pass	pO2 mmHg	104.2	N/A
Na mmol/L	132.1-133.9	Pass	K mmol/L	9.92-10.08	Pass			
K mmol/L	3.97-4.03	Pass	Cl mmol/L	45.6-46.4	Pass			
Cl mmol/L	105.2-106.8	Pass	iCa mmol/L	2.17-2.23	Pass			
iCa mmol/L	1.08-1.12	Pass	Glu mg/dL	197-203	Pass			
Glu mg/dL	79-81	Pass	Lac mmol/L	9.5-10.5	Pass			
Lac mmol/L	1.8-2.2	Pass						

Reference L/N 16231007

	Spec	Pass/Fail
KCl	1.8-2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Approval

QC Supervisor

Title

10/6/2016

Date