



CERTIFICATE OF ANALYSIS

BioPro Flex Cell Density/Viability SA Reagent Pack
Lot Number 16364023

Ref Number 56260
Exp Date 2017-08

W Soln L/N	16348017		C Soln L/N	16348015	
	Spec	Result		Spec	Result
Depro	N/A	Present	3% Detergent pH	N/A	Present
				>9.0	Pass
D Soln L/N	16348023		T Soln L/N	16348013	
	Spec	Result		Spec	Result
Depro	N/A	Present	0.4% Trypan	N/A	Present
%NaOCl	0.43-0.73	Pass	Particle Count	≤ 2.0	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP) , and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH.


Approval

QC Supervisor
Title

2/8/2017
Date