



CERTIFICATE OF ANALYSIS

Linearity Standard Set G Multipack
Lot Number 16005041

Ref Number 56198
Exp Date 2017-12-07
Date of Manufacture 2015-12-07

Level 1 L/N:	15341084
Level 1 Exp Date:	2017-12
Spec	Pass/Fail
Hct %	15-19
Hb g/dL	4.8-6.8
	Pass

Level 1 L/N:	15342013
Level 1 Exp Date:	2017-12
Spec	Pass/Fail
Hct %	58-62
Hb g/dL	18.3-21.3
	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.


Approval

QC Supervisor
Title

1/18/2016
Date