



CERTIFICATE OF ANALYSIS

STP Prime ES Cal Cartridge
Lot Number 16235016

Ref Number 53362
Exp Date 2017-08-16
Date of Manufacture 2016-08-16

Cal A	L/N	16229047	Cal B	L/N	16232004	Reference	L/N	16229044
	Spec	Pass/Fail		Spec	Pass/Fail		Spec	Pass/Fail
Na mmol/L	134.1-135.9	Pass	Na mmol/L	76.2-77.8	Pass	KCl	1.8-2.2	Pass
K mmol/L	4.02-4.08	Pass	K mmol/L	9.8-10.2	Pass			
Cl mmol/L	114.1-115.9	Pass	Cl mmol/L	44.6-45.5	Pass			
iCa mmol/L	0.99-1.01	Pass	iCa mmol/L	1.98-2.02	Pass			
iMg mmol/L	0.46-0.48	Pass	iMg mmol/L	1.49-1.51	Pass			

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Approval

QC Supervisor
Title

8/29/2016
Date