



CERTIFICATE OF ANALYSIS

BioProfile FLEX 2 Chemistry Reference Sensor
Lot Number 18100113

Part Number 57466
Exp Date 2019-10-10
Date of Manufacture 2018-04-10

	Spec*	Pass/Fail
Na Slope	8.7-11.8	Pass
BioPro Ctl L1	70.6-86.2	Pass
BioPro Ctl L3	164.7-201.3	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

*Note: The process for assigning values to Nova Biomedical Quality Control and Linearity Solutions utilizes multiple sources of test data inputs to establish the ranges for each analyte in each lot, which may result in small lot-to-lot variation in insert range values.

A handwritten signature in black ink, appearing to read "Kosmy Bou".

Approval

QC Inspector
Title

4/27/2018
Date