



# **CERTIFICATE OF ANALYSIS**

BioProfile FLEX 2 pH/Gas Reference Sensor  
Lot Number 18103013

Part Number 57467  
Exp Date 2019-10-13  
Date of Manufacture 2018-04-13

|               | Spec*       | Pass/Fail |
|---------------|-------------|-----------|
| pH Slope      | 8.7-11.8    | Pass      |
| BioPro Ctl L1 | 6.636-6.736 | Pass      |
| BioPro Ctl L3 | 7.293-7.393 | Pass      |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

\*Note: The process for assigning values to Nova Biomedical Quality Control and Linearity Solutions utilizes multiple sources of test data inputs to establish the ranges for each analyte in each lot, which may result in small lot-to-lot variation in insert range values.

A handwritten signature in black ink, appearing to read "Kosmy Bau".

Approval

QC Inspector

Title

4/27/2018

Date