



CERTIFICATE OF ANALYSIS

BioProfile FLEX 2 pH/Gas MicroSensor Card LV
Lot Number 18295021

Catalog Number 57510
Exp Date 2019-10-12
Date of Manufacture 2018-10-12

	Slope	Pass/Fail
pH	8.50 - 12.00	Pass
PCO2	-25.00 - 8.00	Pass
PO2	4.00 - 10.00	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

A handwritten signature in black ink, appearing to read "Keo Bou".

Approval

QC Inspector
Title

10/25/2018
Date