



CERTIFICATE OF ANALYSIS

BioProfile FLEX 2 Chemistry Calibrator
and Reagent Cartridge LV

Ref Number 58003

Lot Number 18344019

Exp Date 2019-10-11
Date of Manufacture 2018-10-11

Chemistry Cal Cartridge L/N: 18340003			Exp Date: 2020-01-11		
Cal D1	18284004		Cal D2	18334069	
	Spec	Result		Spec	Result
Gln mmol/L	0.90-1.10	Pass	Gln mmol/L	4.50-5.50	Pass
Glu mmol/L	0.45-0.55	Pass	Glu mmol/L	1.75-2.25	Pass
Gluc g/L	0.90-1.10	Pass	Gluc g/L	4.80-5.80	Pass
Lac g/L	0.90-1.10	Pass	Lac g/L	2.50-3.10	Pass
NH4 mmol/L	1.90-1.98	Pass	NH4 mmol/L	9.93-10.33	Pass
Na+ mmol/L	139.0-143.0	Pass	Na+ mmol/L	65.7-67.7	Pass
K+ mmol/L	3.95-4.07	Pass	K+ mmol/L	9.86-10.18	Pass
Ca++ mmol/L	0.90-0.96	Pass	Ca++ mmol/L	4.50-4.84	Pass

Chemistry Reagent Cartridge L/N: 18325043			Exp Date: 2019-10-11		
DF Soln L/N	18291082		PF Soln L/N	18337013	
	Spec	Pass/Fail		Spec	Result
pH	6.90-7.10	Pass	pH	6.85-7.05	Pass
			R Soln L/N	18274033	
DP Soln L/N	18284013			Spec	Pass/Fail
	Spec	Pass/Fail	KCl mol/L	1.8-2.2	Pass
NaOCl %	0.43-0.73	Pass			

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP) , and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH.

Approval

QC Inspector
Title

12/20/2018
Date