



CERTIFICATE OF ANALYSIS

STP Prime Plus MicroSensor Card
Lot Number 18319044

Ref Number 58643

Exp Date 2019-10-15

Date of Manufacture 2018-10-15

UDI# (01)10385480586433(11)181015(17)191015(10)18319044

	Slope	Pass/Fail
pH	8.5-12.0	Pass
PCO2	0.0-8.0	Pass
PO2	4.0-12.2	Pass
Na	8.5-12.0	Pass
K	8.5-12.0	Pass
Cl	8.5-12.0	Pass
Ca	8.5-14.0	Pass
Mg	8.3-16.6	Pass
Glu	4.0-10.0	Pass
Lac	2.0-11.0	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications.

Approval

QC Inspector
Title

12/5/2018
Date