



CERTIFICATE OF ANALYSIS

Bioprofile pH Sensor
Lot Number 21341043

Part Number 23092
Exp Date 06/30/2023

	<u>Pass/Fail</u>
Slope	Pass
Std A mV	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to test specifications.

Approval

QC Inspector
Title

1/3/2022
Date