

nova
biomedical**CERTIFICATE OF COMPLIANCE**

This product has been manufactured in compliance with CFR 21- Part 820, ISO 13485:2003, and meets Nova Biomedical product specifications.

Description:	StatSensor Creatinine Linearity Kit
Nova Biomedical P/N	44037
	Range
Level 1:	0.5-1.4 mg/dL (44-124 μ mol/L) <u>PASS</u> / FAIL
Level 2:	1.3-2.6 mg/dL (115-230 μ mol/L) <u>PASS</u> / FAIL
Level 3:	2.5-4.5 mg/dL (221-398 μ mol/L) <u>PASS</u> / FAIL
Level 4:	4.9-7.9 mg/dL (433-698 μ mol/L) <u>PASS</u> / FAIL
Level 5:	7.5-12.0 mg/dL (663-1061 μ mol/L) <u>PASS</u> / FAIL
Qty:	325
Lot No.	
Kit:	5508358990
Level 1:	5508340241
Level 2:	5508340242
Level 3:	5508343243
Level 4:	5508344244
Level 5:	5508343245
Expiration Date:	2010-12

m. G. Sullivan
Quality Control12/16/08
Date

Nova Biomedical 200 Prospect St. Waltham, MA 02454-9141