



CERTIFICATE OF ANALYSIS

STP pHox/Basic Cal Cartridge
Lot Number 16063042

Exp Date

Ref Number 23934
2017-08

Cal A	L/N	16054021		Cal B	L/N	16061010		Cal C	L/N	16054023	
		Spec	Pass/Fail			Spec	Pass/Fail			Spec	Pass/Fail
HCO3 mmol/L		20.1-20.9	Pass	HCO3 mmol/L		40.2-41.8	Pass	pH @ 37C		7.342-7.348	Pass
Na mmol/L		139.3-141.3	Pass								
Cal D	L/N	16054024		Reference	L/N	16054025					
		Spec	Pass/Fail			Spec	Pass/Fail				
pH @ 37C		6.792-6.798	Pass	KCl		1.8-2.2	Pass				
Na mmol/L		71.5-72.5	Pass								

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Approval

QC Supervisor

Title

3/15/2016

Date