



CERTIFICATE OF ANALYSIS

STP Prime Plus MicroSensor Card Bun, Creatinine

Ref Number 57821

Lot Number 18292034

Exp Date 2019-10-15

Date of Manufacture 2018-10-15

UDI# (01)10385480578216(11)181015(17)191015(10)18292034

	Slope Spec	Pass/Fail
Bun	8.0 - 16.0	Pass
Creatinine	2.0 - 15.0	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications.

A handwritten signature in purple ink that reads "Keo Bou".

Approval

QC Inspector
Title

1/14/2019
Date