

OI-922

Page 33 of 66

Rev. BH

Vialed and Packaged Test Strip Release

Test Strips

nova
biomedical

CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	N/A
Description:	TEST STRIPS, STATSTRIP™ GLUCOSE, PACKAGED 1800 COUNT
Catalog Number:	42214
Nova Biomedical P/N	51493
Target Range (mg/dL) Mean MPE (%) CV (%)	
60-75 ≤ 15 ≤ 6.5	PASS / FAIL
120-160 ≤ 15 ≤ 5.5	PASS / FAIL
250-290 ≤ 15 ≤ 5.0	PASS / FAIL
380-420 ≤ 15 ≤ 5.0	PASS / FAIL
500-540 ≤ 15 ≤ 5.0	PASS / FAIL
Combined MPE (%) ≤ 10%	PASS / FAIL
Qty:	2,520
Lot No.	0322356309
Mfg. Date:	2022-12-22
Expiration Date:	2025-06-22
Unique Device Identifier (UDI):	(01)20385480422141 (11)221222 (17)250622 (10)0322356309

Max Brune
Quality Control Manager

01/25/2023
Date

Nova Biomedical Corporation

200 Prospect Street

Waltham, MA 02454