



CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with CFR 21- Part 820, ISO 13485:2012, and meets Nova Biomedical product specifications.

Customer:	N/A
Description:	TEST STRIPS, STATSTRIP™ GLUCOSE, PACKAGED 1800 COUNT
Catalog Number:	42214
Nova Biomedical P/N	42214
Target Range (mg/dL) Mean MPE (%) CV (%) 60-75 ≤ 15 ≤ 6.5 PASS / FAIL 120-160 ≤ 15 ≤ 5.5 PASS / FAIL 250-290 ≤ 15 ≤ 5.0 PASS / FAIL 380-420 ≤ 15 ≤ 5.0 PASS / FAIL 500-540 ≤ 15 ≤ 5.0 PASS / FAIL Combined MPE (%) ≤ 10% PASS / FAIL	
Qty:	1458, 3/14 Singles
Lot No.	0317283249
Mfg. Date:	2017/10/10
Expiration Date:	2019/10/10

 Quality Control Manager

11/2/17

 Date

Nova Biomedical Corporation

200 Prospect Street Waltham, MA 02454