

OI-922

Page 55 of 66

Rev. BH

Vialed and Packaged Test Strip Release

Test Strips

nova
biomedical

CERTIFICATE OF COMPLIANCE Page 1/2

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIPRO DIAGNOSTICS INC.
Description:	TEST STRIPS, NIPRO STATSTRIP™ GLUCOSE, PACKAGED
Catalog Number:	48880
Nova Biomedical P/N	48880
Target Range (mg/dL) Mean MPE (%) CV (%)	
60-75 ≤ 15 ≤ 6.5	PASS / FAIL
120-160 ≤ 15 ≤ 5.5	PASS / FAIL
250-290 ≤ 15 ≤ 5.0	PASS / FAIL
380-420 ≤ 15 ≤ 5.0	PASS / FAIL
500-540 ≤ 15 ≤ 5.0	PASS / FAIL
Combined MPE (%) ≤ 10%	PASS / FAIL
Qty:	104
Lot No.	2122318309
Mfg. Date:	2022.11.14
Expiration Date:	2025.05.31

Marc Broome
Quality Control Manager

11/28/2022
Date

Nova Biomedical Corporation

200 Prospect Street

Waltham, MA 02454

Released Confidential

OI-922 Page 56 of 66 Rev. BH Vial and Packaged Test Strip Release Test Strips

nova
biomedical

ADDENDUM TO CERTIFICATE OF COMPLIANCE Page 2/2

RELEASE TEST RESULTS

Customer:	NIPRO DIAGNOSTICS INC.
Description:	TEST STRIPS, NIPRO STATSTRIP™ GLUCOSE
Catalog Number:	48880
Nova Biomedical P/N	48880
Lot No.	2122318309

(Spiked whole blood) Target Range:60-75			
Mean (mg/dL) N=72	CV%	YSI (reference, mg/dL)	Mean MPE%
68	4.4 ≤6.5	70.8	5.0 ≤15

(Spiked whole blood) Target Range:120-160			
Mean (mg/dL) N=72	CV%	YSI (reference, mg/dL)	Mean MPE%
149	4.6 ≤5.5	148.8	3.4 ≤15

(Spiked whole blood) Target Range:250-290			
Mean (mg/dL) N=72	CV%	YSI (reference, mg/dL)	Mean MPE%
264	3.6 ≤5.0	274.3	4.3 ≤15

(Spiked whole blood) Target Range:380-420			
Mean (mg/dL) N=72	CV%	YSI (reference, mg/dL)	Mean MPE%
386	3.6 ≤5.0	406.3	5.0 ≤15

(Spiked whole blood) Target Range:500-540			
Mean (mg/dL) N=72	CV%	YSI (reference, mg/dL)	Mean MPE%
488	3.4 ≤5.0	522.8	6.6 ≤15

Marc Broome
Quality Control Manager

11/28/2022
Date

Nova Biomedical Corporation 200 Prospect Street Waltham, MA 02454