

OI-922

Page 55 of 66

Rev. BH

Vialed and Packaged Test Strip Release

Test Strips

nova
biomedical

CERTIFICATE OF COMPLIANCE Page 1/2

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIPRO DIAGNOSTICS INC.
Description:	TEST STRIPS, NIPRO STATSTRIP™ GLUCOSE, PACKAGED
Catalog Number:	48880
Nova Biomedical P/N	48880
Target Range (mg/dL) Mean MPE (%) CV (%)	
60-75 ≤ 15 ≤ 6.5	PASS / FAIL
120-160 ≤ 15 ≤ 5.5	PASS / FAIL
250-290 ≤ 15 ≤ 5.0	PASS / FAIL
380-420 ≤ 15 ≤ 5.0	PASS / FAIL
500-540 ≤ 15 ≤ 5.0	PASS / FAIL
Combined MPE (%) ≤ 10%	PASS / FAIL
Qty:	105
Lot No.	2123016309
Mfg. Date:	2023.01.16
Expiration Date:	2025.07.31

Marc Brosme
Quality Control Manager

1/19/2023
Date

Nova Biomedical Corporation

200 Prospect Street Waltham, MA 02454

OI-922 Page 56 of 66 Rev. BH Viald and Packaged Test Strip Release Test Strips

nova
biomedical

ADDENDUM TO CERTIFICATE OF COMPLIANCE Page 2/2

RELEASE TEST RESULTS

Customer:	NIPRO DIAGNOSTICS INC.
Description:	TEST STRIPS, NIPRO STATSTRIP™ GLUCOSE
Catalog Number:	48880
Nova Biomedical P/N	48880
Lot No.	2123016309

(Spiked whole blood) Target Range:60-75			
Mean (mg/dL) N=72	CV%	YSI (reference, mg/dL))	Mean MPE%
	≤6.5		≤15
70	5.7	69.1	4.8

(Spiked whole blood) Target Range:120-160			
Mean (mg/dL) N=72	CV%	YSI (reference, mg/dL))	Mean MPE%
	≤5.5		≤15
143	3.6	146.5	3.2

(Spiked whole blood) Target Range:250-290			
Mean (mg/dL) N=72	CV%	YSI (reference, mg/dL))	Mean MPE%
	≤5.0		≤15
266	3.0	269.8	2.6

(Spiked whole blood) Target Range:380-420			
Mean (mg/dL) N=72	CV%	YSI (reference, mg/dL))	Mean MPE%
	≤5.0		≤15
402	3.5	410.3	3.3

(Spiked whole blood) Target Range:500-540			
Mean (mg/dL) N=72	CV%	YSI (reference, mg/dL))	Mean MPE%
	≤5.0		≤15
508	3.0	515.0	2.6

Quality Control Manager

Date

Nova Biomedical Corporation

200 Prospect Street Waltham, MA 02454