



CERTIFICATE OF ANALYSIS

STP pHox /Basic/Plus/L/C/M pCO2 Sensor
Lot Number 21347044

Ref Number 21524

Exp Date 2023-12-13

Date of Manufacture 2021-12-13

UDI (01)00385480215244(11)211213(17)231213(10)21347044

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.50 - 11.30	Pass
Std A mV	0.00 +/- 150.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector
Title

1/7/2022
Date