nova biomedical

CERTIFICATE OF ANALYSIS

STP pHOx /Basic/Plus/L/C/M pCO2 Sensor — Ref Number 21524

Lot Number 22179046 — Exp Date 2024-06-28

Date of Manufacture 2022-06-28

UDI (01)00385480215244(11)220628(17)240628(10)22179046

	Spec	Pass/Fail
Slope	9.50 - 11.30	Pass
Std A mV	0.00 +/- 150.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector
Title

7/13/2022
Date