



**CERTIFICATE OF ANALYSIS**

STP pHox /Basic/Plus/L/C/M pO2 Sensor  
Lot Number 22269024

Ref Number 21521  
Exp Date 2024-09-26

Date of Manufacture 2022-09-26

UDI# (01)00385480215213(11)220926(17)240926(10)22269024

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	-3.00 to -6.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector  
Title

10/12/2022  
Date