



CERTIFICATE OF ANALYSIS

Bioprofile 100 Ammonium Sensor
Lot Number 15260029

Part Number 24451
Exp Date 9/30/2017

	Spec	Pass/Fail
Slope	10.5-13.0	Pass
Std J mV	≥ -100	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

A handwritten signature in dark ink, appearing to read "Joyce Ansara".

Approval

QC Supervisor

Title

10/7/2015

Date