



CERTIFICATE OF ANALYSIS

Bioprofile 100/300 pH Sensor
Lot Number 21237038

Part Number 24453
Exp Date 02/28/2023

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.2 - 11.2	Pass
Std A mV	0 ± 50	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

9/3/2021