



**CERTIFICATE OF ANALYSIS**

Bioprofile 100/400/100 Plus Reference Electrode  
Lot Number 21197003

Part Number 24454  
Exp Date 01/31/2023

	<u>Spec*</u>	<u>Pass/Fail</u>
Na Slope	≥ 8.9	Pass
Std C mV	-80.00 to +80.00	Pass
BioPro Ctl L2	124.7 - 138.7	Pass
K Slope	≥ 9.8	Pass
Std A mV	-50.00 to +50.00	Pass
BioPro Ctl L2	3.65 - 4.35	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

\*Note: The process for assigning values to Nova Biomedical Quality Control and Linearity Solutions utilizes multiple sources of

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

8/10/2021  
Date