



CERTIFICATE OF ANALYSIS

Bioprofile 400/Basic/100 Plus Glutamine Membrane
Lot Number 21272026

Part Number 24456
Exp Date 2022-09

	<u>Spec*</u>	<u>Mean Result</u>
Slope	10.0 - 60.0	23.3
Biopro Ctl L4	0.82 - 1.32	1.15
Biopro Ctl L5	4.54 - 6.04	5.22

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi.M

Approval

QC Inspector
Title

10/15/2021
Date