



CERTIFICATE OF ANALYSIS

Bioprofile 400/Basic/100 Plus Lactate Membrane
Lot Number 23047063

Part Number 24457
Exp Date 2024-02

	<u>Spec*</u>	<u>Mean Result</u>
Slope	10.0 - 60.0	36.9
Biopro Ctl L1	0.89 - 1.39	1.07
Biopro Ctl L2	2.55 - 3.45	3.20
Biopro Ctl L3	4.25 - 5.75	4.92

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in dark ink, appearing to read "Hermann Aichta", written in a cursive style.

Approval

QC Inspector
Title

2/24/2023
Date