



**CERTIFICATE OF ANALYSIS**

STP pHox Plus Chloride Sensor  
Lot Number 22249033

Part Number 27812  
Exp Date 2023-12-06  
Date of Manufacture 2022-09-06  
UDI# (01)10385480278123(11)220906(17)231206(10)22249033

	<u>Spec*</u>	<u>Pass/Fail</u>
Slope	≥ 9.00	Pass
Std A mV	55.00 - 95.00	Pass
pHOx Ctl L1	120.7 - 129.7	Pass
pHOx Ctl L3	79.7 - 89.7	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

10/13/2022  
Date