



## CERTIFICATE OF ANALYSIS

STP pHox/pHOx Basic/BiopHOx Air Detector  
Lot Number 22060082

Ref Number 21513  
Date of Manufacture 2022-03-01

|        | <u>Spec</u> | <u>Pass/Fail</u> |
|--------|-------------|------------------|
| AD1 mV | 1400 - 2600 | Pass             |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

A handwritten signature in black ink, appearing to read "J. P. Cant", with a stylized flourish at the end.

Approval

QC Inspector  
Title

3/10/2022  
Date