



CERTIFICATE OF ANALYSIS

Nova Cleaning Solution
Lot Number 22040031

Ref Number 11272
Exp Date 2023-02-09
Date of Manufacture 2022-02-09
UDI# (01)00385480112727(11)220209(17)230209(10)22040031

| | <u>Spec</u> | <u>Pass/Fail</u> |
|-------|--------------|------------------|
| NaOCl | 1.21 - 1.51% | Pass |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector
Title

2/17/2022
Date