



CERTIFICATE OF ANALYSIS

Bioprofile 300 Glucose Membrane
Lot Number 21327016

Part Number 33255
Exp Date 2023-02

	<u>Spec*</u>	<u>Mean Result</u>
Slope	12.0 - 60.0	43.2
Biopro Ctl L1	0.75 - 1.25	1.06
Biopro Ctl L3	12.75 - 17.25	14.31

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

12/14/2021
Date