



**CERTIFICATE OF ANALYSIS**

Stat Profile® pHox Ultra™ / CCX™ SO2 Sensor  
Lot Number 21306012

Part Number 48591  
Exp Date 2024-11-02  
Date of manufacture 2021-11-02  
UDI#(01)00385480485913(11)211102(17)241102(10)21306012

	<u>Spec</u>	<u>Mean Result</u>	<u>Pass/Fail</u>
Slope	7.00 - 14.00	8.95	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

11/11/2021  
Date