



CERTIFICATE OF ANALYSIS

StatStrip Glucose Control Solution Level 1
LifeScan Reference 02315401
Lot Number 0422007301

Date of Manufacture 2022-01-07
Exp Date 2024-07-07

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	57 - 65	62

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

2/2/2022
Date